National Functional Guidelines Report # 3

11:30 Tue, Oct 28, 2008

Lab MITKEM (Mitkem Corporation)SDG Y4FR4Case 37947Contract EPW05030Region 9DDTID 65388SOW SOM01.2Data Review Results

Initial Calibration

Initial Calibration	VOA_LOW_MED
VC15	The following volatile samples are associated with an initial calibration with relative response factors (RRFs) outside criteria. Detected compounds are qualified J. Nondetected compounds are qualified R.
	VBLK5Q, VHBLK5Q, Y4FR4, Y4FS2, Y4FW0
	1,4-Dioxane VSTD0055Q
	Isopropanol VSTD0055Q, VSTD0105Q, VSTD0505Q, VSTD1005Q, VSTD2005Q
Initial Calibration	VOA_LOW_MED
VC6	The following volatile samples are associated with an initial calibration percent relative standard deviation (%RSD) outside criteria. Detected compounds are qualified J. Nondetected compounds are not qualified. Use professional judgement to qualify non-detected compounds.
	VBLK5Q, VHBLK5Q, Y4FR4, Y4FS2, Y4FW0
	Bromomethane BFB50
	VBLK5Q, VHBLK5Q, Y4FR4, Y4FS2, Y4FW0